

REPORT DOCUMENTATION PAGE

AFRL-SR-BL-TR-01-

Public reporting burden for this collection of information is estimated to average 1 hour per response, including the time for reviewing instructions, data needed, and completing and reviewing this collection of information. Send comments regarding this burden estimate or any other aspect of this burden to Department of Defense, Washington Headquarters Services, Directorate for Information Operations and Reports (0704-0188), 4302. Respondents should be aware that notwithstanding any other provision of law, no person shall be subject to any penalty for failing to provide a valid OMB control number. PLEASE DO NOT RETURN YOUR FORM TO THE ABOVE ADDRESS.

0321

1. REPORT DATE (DD-MM-YYYY) 08/02/2001		2. REPORT TYPE Final Technical Report		3. DATES COVERED (From - To) 01/09/1996 - 31/03/2000	
4. TITLE AND SUBTITLE Gas-Phase Reactions of Negative Ions at Hyperthermal Energies				5a. CONTRACT NUMBER	
				5b. GRANT NUMBER F49620-96-0409	
				5c. PROGRAM ELEMENT NUMBER	
6. AUTHOR(S) Peter M. Hierl				5d. PROJECT NUMBER	
				5e. TASK NUMBER	
				5f. WORK UNIT NUMBER	
7. PERFORMING ORGANIZATION NAME(S) AND ADDRESS(ES) The University of Kansas The University of Kansas Research Support & Grants Administration 221 Strong Hall Lawrence, KS 66045				8. PERFORMING ORGANIZATION REPORT NUMBER	
9. SPONSORING / MONITORING AGENCY NAME(S) AND ADDRESS(ES) Air Force Office of Scientific Research 801 N. Randolph St., Rm. 732 Arlington, VA 22203-1977				10. SPONSOR/MONITOR'S ACRONYM(S) AFOSR/NL	
				11. SPONSOR/MONITOR'S REPORT NUMBER(S)	
12. DISTRIBUTION / AVAILABILITY STATEMENT DISTRIBUTION STATEMENT A Approved for Public Release Distribution Unlimited					
13. SUPPLEMENTARY NOTES					
14. ABSTRACT A tandem mass spectrometer incorporating radio-frequency octopole ion guides has been designed, constructed, and tested. This instrument is capable of measuring absolute integral cross sections for gas-phase ion molecule reactions as functions of the reactant's relative translation energy from thermal to hyperthermal energy (0.03 - 20 eV c.m.), and of the neutral reactant's temperature from 298 - 1000 K. A series of calibration studies have been carried out to verify the operating characteristics of this new instrument.					
15. SUBJECT TERMS					
16. SECURITY CLASSIFICATION OF:			17. LIMITATION OF ABSTRACT UU	18. NUMBER OF PAGES 3	19a. NAME OF RESPONSIBLE PERSON Barbara Armbrister
a. REPORT U	b. ABSTRACT U	c. THIS PAGE U			19b. TELEPHONE NUMBER (include area code) (785) 864-3444

20010521 179

Final Report
AFOSR DEPSCoR Grant
#F49620-96-1-0409
PI: Peter M. Hierl

1. Objectives

The objectives of this project were (1) to design, build, and test a new instrument for the measurement of absolute integral cross sections of ion-molecule reactions as functions of reactant translational energy and of neutral reactant temperature; and (2) to perform such measurements on several electron-transfer and proton-transfer reactions relevant to the high-temperature plasma surrounding vehicles re-entering the earth's atmosphere.

2. Status of Efforts

A tandem mass spectrometer for the measurement of absolute integral cross sections of ion-molecule reactions at high temperatures has been designed, built, and successfully tested in a series of calibration experiments. Work has begun (and is still in progress) on using this instrument to study selected ion-molecule reactions relevant to the high-temperature plasma surrounding vehicles re-entering the earth's atmosphere.

3. Accomplishments

The major accomplishment has been the design and construction of a tandem mass spectrometer which incorporates octopole ion guides to transmit the reactant ion beam through a high-temperature collision cell containing the neutral reactant, and to transmit reactant and product ions to the detector. The details of this device are given below:

The instrument consists of three differentially-pumped vacuum chambers: an ion source, a reaction region, and an analyzer region. These chambers are pumped by an 1800 L/s diffusion pump, a 5000 L/s diffusion pump, and a 550 L/s turbomolecular pump, respectively. Background pressures in these three regions are $\sim 5 \times 10^{-8}$ torr, and operating pressures are less than 2×10^{-7} torr.

The reactant ions are generated by electron impact on a precursor gas in a low-pressure source, extracted and focused into a beam which passes at moderate energy (20-100 eV) through a small magnetic sector mass spectrometer. The emerging mass-selected beam is focused by a second

electrostatic lens system into a nearly monoenergetic beam whose energy can be varied from ~0.1-100 eV.

This beam is then injected into the first of two coupled octopole ion guides (i.e., eight parallel rods in a circular array on which opposite phases of a radiofrequency (rf) voltage are applied, thereby creating a confining electric field which guides the ions along the axis of the rods). While traversing the first ion guide (30 cm long), the ions pass through a variable-temperature (298-1000 K) collision cell containing the neutral reactant gas, the pressure of which (typically, ~0.2 mtorr) is measured with a capacitance manometer.

Upon exiting the first octopole ion guide, product ions and transmitted primary ions immediately enter a second octopole ion guide (60 cm long) which is used to measure the translational energy of the ions. To date, we have used a continuous ion beam and performed the energy analysis by using the second octopole as a retarding potential analyzer. It is also possible, however, to pulse the ion beam and use the second octopole as a time-of-flight energy analyzer.

Ions exiting the second ion guide are focused onto the entrance of a quadrupole mass spectrometer for identification. Detection utilizes an electron multiplier and standard pulse counting techniques. The instrument is interfaced with a desktop computer for control of the experimental parameters and for data acquisition.

The operating characteristics of the instrument have been determined in a series of calibration experiments. The instrument was found capable of producing intense ($0.5\text{--}2 \times 10^6$ cps), nearly monoenergetic (FWHM = 0.15-0.20 eV) mass selected ion beams over the range of laboratory energies 0.1-100 eV. Integral cross sections were measured at room temperature for a number of well-known ion-molecule reactions; our results were found to be in excellent agreement with data previously published by other authors. For the reaction $\text{Ar}^+ + \text{D}_2 \rightarrow \text{ArD}^+ + \text{D}$, measurements were also made at temperatures up to 900 K; the results, corrected for the effects of thermal transpiration, confirmed the proper operation of the instrument at elevated temperatures. In another set of experiments, cross sections were measured for several endoergic ion-molecule reactions (e.g., $\text{Ne}^+ + \text{CO} \rightarrow \text{Ne} + \text{C}^+ + \text{O}$). The translational energy thresholds observed were in excellent agreement with the known thermodynamic thresholds, indicating both the accuracy of our energy measurements and the utility of this technique for determining such values.

4. Personnel Supported:

Yousef Basir (postdoctoral student, 18 months)
Nadya Galeva (postdoctoral student, 6 months)

5. Publications: none at this time

6. Interactions

a. Participation:

During the period covered in this report, I have been a regular participant at the Molecular Dynamics Contractor's Reviews sponsored by AFOSR, and at Gordon Research Conferences on gas-phase ion chemistry.

b. Consultative/advisory functions

During this project I have visited and consulted with several colleagues (Rainer Dressler, Robert Morris, Albert Viggiano, Skip Williams, and others) at the Air Force Research Laboratory, Hanscom AFB, MA. The subjects discussed included design features of the guided ion beam tandem mass spectrometer, specific ion chemistry relating to the current project, and the possibility of continued collaboration in the future (e.g., ion-enhanced combustion of hydrocarbon fuels).

I have also visited Professors Scott Anderson and Peter Armentrout (University of Utah), and Professor Richard Zare (Stanford University) to discuss the design of this instrument, and have received visits from Rainer Dressler (AFRL) and Professors Kent Erwin (University of Nevada at Reno) and Dieter Gerlich (Technical University of Chemnitz, Germany) for similar discussions.

7. New discoveries: none at this point

8. Honors/Awards: none at this point